
## **BULK CURRENT INJECTION**

The F-140-HV is used for BCI testing in accordance with DO-160, section 20, as well as Pulse Inductive Damped Sinusoid testing in accordance with DO-160, section 22.





## **Specifications**

Frequency: 100 kHz – 400 MHz

Inner Diameter: 40 mm
Outer Diameter: 127 mm
Height: 70 mm
Peak Pulse Rating: 5 kV
RF Connector: HN Female

Secondary Power Current Rating: DC - 400 Hz
• Differential Mode - Single Conductor: 100 Amps
• Common Mode -Total Cable Bundle: 500 Amps

## **Accessories**

FCC-BCICF-1-HV • Calibration Fixture Bandwidth: 10 kHz – 400 MHz

Specifications subject to change without notice



19220 Normandie Ave., Unit B Torrance, CA 90502

Phone: 310-303-3300 Email: sales@fischercc.com